



Designation: D1613 – 17

Standard Test Method for Acidity in Volatile Solvents and Chemical Intermediates Used in Paint, Varnish, Lacquer, and Related Products¹

This standard is issued under the fixed designation D1613; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reappraisal.

This standard has been approved for use by agencies of the U.S. Department of Defense.

1. Scope*

1.1 This test method covers the determination of total acidity as acetic acid, in concentrations below 0.05 %, in organic compounds and hydrocarbon mixtures used in paint, varnish, and lacquer solvents and diluents. It is known to be applicable to such mixtures as low molecular weight saturated and unsaturated alcohols, ketones, ethers, esters, hydrocarbon diluents, naphtha, and other light distillate petroleum fractions.

1.2 For purposes of determining conformance of an observed value or a calculated value using this test method to relevant specifications, test result(s) shall be rounded off “to the nearest unit” in the last right-hand digit used in expressing the specification limit, in accordance with the rounding-off method of Practice E29.

1.3 The values stated in SI units are to be regarded as the standard. The values given in parentheses are for information only.

1.4 For specific hazard information and guidance consult supplier’s Safety Data Sheet.

1.5 *This standard does not purport to address the safety concerns, if any, associated with its use. It is the responsibility of whoever uses this standard to consult and establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

¹ This test method is under the jurisdiction of ASTM Committee D01 on Paint and Related Coatings, Materials, and Applications and is the direct responsibility of Subcommittee D01.35 on Solvents, Plasticizers, and Chemical Intermediates.

Current edition approved Feb. 1, 2017. Published February 2017. Originally approved in 1964. Last previous edition approved in 2012 as D1613 – 06 (2012). DOI: 10.1520/D1613-17.

2. Referenced Documents

2.1 *ASTM Standards:*²

D770 Specification for Isopropyl Alcohol

D1193 Specification for Reagent Water

D4806 Specification for Denatured Fuel Ethanol for Blending with Gasolines for Use as Automotive Spark-Ignition Engine Fuel

D7795 Test Method for Acidity in Ethanol and Ethanol Blends by Titration

E29 Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications

E200 Practice for Preparation, Standardization, and Storage of Standard and Reagent Solutions for Chemical Analysis

3. Summary of Test Method

3.1 The specimen is mixed with either an equal volume of water or an equal volume of alcohol, and titrated with aqueous sodium hydroxide solution to the phenolphthalein end point.

4. Significance and Use

4.1 This test method is useful for determining low levels of acidity, below 0.05 %, in organic compounds and hydrocarbon mixtures. The total acidity is calculated as acetic acid or milligrams of sodium hydroxide per gram of sample.

4.2 Acidity may be present as a result of contamination, decomposition during storage or distribution, or manufacture. This test method may be used in assessing compliance with a specification.

5. Apparatus

5.1 *Buret*, 10-mL, graduated in 0.05-mL subdivisions.

5.2 *Erlenmeyer Flask*, 250-mL capacity.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard’s Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard

6. Purity of Reagents

6.1 Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.³ Other grades may be used provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

6.2 Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Type IV of Specification D1193.

7. Reagents

7.1 *Alcohols*, refined, ethyl or isopropyl.

NOTE 1—Isopropyl alcohol (99 % grade) conforming to Specification D770, or 190 proof ethyl alcohol conforming to formula No. 3A of the U.S. Bureau of Alcohol, Tobacco and Firearms is suitable for use as the solvent. The use of methyl alcohol is not recommended.

7.2 *Phenolphthalein Indicator Solution (10 g/L)*—Dissolve 1 g of phenolphthalein in ethyl or isopropyl alcohol (see Note 1) and dilute to 100 mL with the alcohol.

7.3 *Sodium Hydroxide, Standard Solution (0.05 N)*—Prepare and standardize a 0.05 N sodium hydroxide (NaOH) solution (Note 2) in accordance with the Preparation and Standardization of Solutions, Precision and Bias, Preparation of 50 % of NaOH Solution and of Standard Solutions, and Standardization sections of Practice E200.

NOTE 2—Alternatively, KOH solution may be used.

8. Procedure

8.1 Measure into a 250-mL Erlenmeyer flask 50 mL of water, if the sample is completely water-soluble, or 50 mL of alcohol, if the sample is not completely water-soluble.

8.2 Add 0.5 mL of phenolphthalein indicator solution. Titrate the water or alcohol with 0.05 N NaOH solution to the first perceptible pink color.

8.3 Pipet 50 mL of the sample into the flask. Titrate with the 0.05 N NaOH solution to the same first perceptible pink color originally obtained.

8.3.1 If the sample is fuel ethanol or denatured fuel ethanol, as defined in Specification D4806, purge the sample for 2 min with 400 mL/min of nitrogen prior to pipetting (see Test Method D7795 for additional information on purging the sample).

NOTE 3—It is well known that carbon dioxide is highly soluble in ethanol solutions. Carbon dioxide can be dissolved into the ethanol solution during production (especially fermentation processes), during transportation or during laboratory analysis. The dissolved carbon dioxide is a known interference for this test method creating incorrect elevated

³ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

acidity values. If the dissolved carbon dioxide is not removed from a sample, erroneously high results can be experienced which can exceed the specification limit. It is recommended that Test Method D7795 be used for the determination of acidity in fuel ethanol and denatured fuel ethanol; furthermore, as is codified in Specification D4806, Test Method D7795 should be considered the referee method:

Section 8.4.1 from Specification D4806: “Dissolved carbon dioxide is a known interference and can cause a false high reading when using Test Method D1613. In the absence of dissolved CO₂, Test Method D1613 is an acceptable method. If a sample is known to have dissolved CO₂ or if dissolved CO₂ can be present, Test Method D7795 is the preferred method. In cases of differing results between the two test methods, Test Method D7795 shall be the referee method.”

9. Calculations

9.1 Calculate the acidity of the sample as follows:

$$\text{Acidity as acetic acid, weight \%} = (VN \times 0.12)/D \quad (1)$$

or,

$$\text{Acidity as mg KOH per g of sample} = (VN \times 1.12)/D \quad (2)$$

where:

V = NaOH solution required for titration of the sample, mL,

N = normality of the NaOH solution, and

D = density of specimen in g/mL.

10. Report

10.1 Report the percent of acetic acid to the nearest 0.0001 %. Duplicate runs that agree within 0.0005 %, absolute, are acceptable for averaging (95 % confidence level).

11. Precision and Bias⁴

11.1 *Precision*:

11.1.1 The following criteria should be used for judging the acceptability of results at the 95 % confidence level:

11.1.1.1 *Repeatability*—The normal range between two results, each the mean of duplicate determinations, obtained by the same analyst on different days, is estimated to be 0.0003 %, absolute. Two such values should be considered suspect if they differ by more than 0.0008 %, absolute.

11.1.1.2 *Reproducibility*—The normal range between two results, each the mean of duplicate determinations obtained by analysts in different laboratories, is estimated to be 0.0005 %, absolute. Two such values should be considered suspect if they differ by more than 0.0014 %, absolute.

NOTE 4—The above precision estimates are based on an interlaboratory study on two samples each of *n*-butyl acetate, *n*-butyl alcohol, and methyl ethyl ketone containing 0.0058, 0.0112, 0.0007, 0.0046, 0.0026, and 0.0067 % acetic acid, respectively. Each of four laboratories analyzed all six samples, with two analysts in each laboratory performing duplicate determinations using both 99 % isopropyl alcohol and formula 3A ethanol as solvents, and repeating on a second day, for a total of 384 determinations.

11.2 *Bias*—The bias of this test method has not been determined because there is no available material with an accepted reference value.

⁴ Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D01-1041. Contact ASTM Customer Service at service@astm.org.

12. Keywords

12.1 acidity; solvents; total acidity as acetic acid

SUMMARY OF CHANGES

Committee D01 has identified the location of selected changes to this standard since the last issue (D1613–06(2012)) that may impact the use of this standard. (Approved February 1, 2017.)

(1) Added D4806 Specification for Denatured Fuel Ethanol for Blending with Gasolines for Use as Automotive Spark-Ignition Engine Fuel as a referenced document in Section 2.
(2) Added D7795 Standard Test Method for Acidity in Ethanol and Ethanol Blends by Titration as a referenced document in Section 2.

(3) Added new section 8.3.1 to give specific direction for using Test Method D1613 on fuel ethanol or denatured fuel ethanol.
(4) Updated Note 3 to give specific directions for the purge conditions required when using Test Method D1613 on fuel ethanol or denatured fuel ethanol.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>